

Minutes of Pre-Bid Conference held on 25 August 2021 @ 02.30 P.M
Through MS Teams

Open Tender No. 4/5(439)21-Purdated 12 August 2021
Differential Scanning Calorimeter

The following attended the conference.

From Technical & Purchase Committee

Dr. Vipin Hallan, SPS	Chairman
Dr. Arun Kumar, SS	IO-Member
Dr. Upendra Sharma, PS	Member
Shri. Bijan, STO	Member
Shri Amarjeet, COA	Member
Shri Yashpal, FAO	Member
Shri Yashpal, SPO	Member-Convener

From Bidders:

- 1) M/s. Waters India Pvt. Ltd.
- 2) M/s. Mettler Toledo India Pvt. Ltd.
- 3) M/s. Inkar Instruments Pvt. Ltd.


DELIBERATIONS:

The TPC had given the clarifications to the issues raised by M/s. Waters, and M/s, Inkar Instruments vide their e-mails. Further the TPC also modified certain specifications to make it more clear to the bidders. The clarifications given and specifications modified/added are stated in Annexure A.


IO-Member



Member


27/08/2021
Member


AO 27/8


FAO 27/8/21


SPO 27/8/21

Chairman
Appeared online

27/08/21

ANNEXURE- A

MODIFIED SPECIFICATIONS		
	Existing specifications	Revised specifications
Principle of operation:	Heat Flux	Sinusoidal temperature modulated heat flux
Temp. accuracy :	± 0.2-0.5 °C or better	± 0.2 °C/K or better
Press and die set for preparation of pellets:	Required	Appropriate/ compatible sealing press which can seal/ crimp above mentioned Aluminium crucibles.

Note:

- An updated list of users in India where the quoted model/higher model is already installed should be provided with their contact details.

Signature

IO-Member

Signature

Member

Signature
27/08/2024

Member

Signature
27/8

AO

Signature

FAO

Signature
27/8/24

ASPO

Chairman

Signature
Approved Online
27/08/21